

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2638

Name: RITTER 1  
Sterilizer ID: V2178198  
12/05/2022 01:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.6 F 0.0  
2:00 176.3 F 2.4  
4:00 175.8 F 6.1  
6:00 201.1 F 10.9  
8:00 227.1 F 6.1  
10:00 237.3 F 11.6  
12:00 235.7 F 8.5  
14:00 241.3 F 10.6  
16:00 246.1 F 13.1  
18:00 258.1 F 19.4  
20:00 268.6 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 271.9 F 29.0  
2:00 272.4 F 28.8  
3:00 273.7 F 29.5  
4:00 273.0 F 29.4

Min 270.9 F 28.5  
Max 273.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
5:00 235.8 F 0.0  
10:00 224.6 F 0.0  
15:00 229.4 F 0.0  
20:00 229.2 F 0.0  
25:00 226.6 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:03  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:59:34  
CYCLE COMPLETE

12/05/2022 02:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_