
Midmark M11 - v1.0.5
Total Cycles: 1131

Name: RITTER2
Sterilizer ID: V2313228
05/27/2022 12:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.9 F 0.0
2:00 180.4 F 3.8
4:00 187.6 F 8.4
6:00 208.4 F 13.5
8:00 224.1 F 8.2
10:00 236.9 F 13.6
12:00 238.5 F 11.2
14:00 237.2 F 9.9
16:00 251.2 F 16.5
18:00 262.8 F 23.3
20:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 29.8
2:00 271.5 F 29.0
3:00 272.3 F 29.9
4:00 272.4 F 29.6

Min 271.2 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 216.3 F 0.0
10:00 231.8 F 0.0
15:00 232.7 F 0.0
20:00 225.8 F 0.0
25:00 223.0 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:02
CYCLE COMPLETE

05/27/2022 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____