
Midmark M11 - v1.0.0
Total Cycles: 5087

Name: RITTER 1
Sterilizer ID: V2178198
11/13/2024 05:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 155.8 F 0.3
4:00 173.3 F 0.9
6:00 197.2 F 4.1
8:00 215.8 F 8.4
10:00 226.1 F 13.3
12:00 231.4 F 9.0
14:00 238.2 F 9.7
16:00 244.2 F 13.0
18:00 243.7 F 12.0
20:00 256.9 F 19.0
22:00 268.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.0 F 28.6
2:00 272.5 F 29.0
3:00 271.6 F 28.2
4:00 272.3 F 28.9

Min 271.4 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.9 F 0.0
5:00 225.0 F 0.0
10:00 212.9 F 0.0
15:00 219.3 F 0.0
20:00 218.2 F 0.0
25:00 221.1 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:09
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:43
CYCLE COMPLETE

11/13/2024 06:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____