
Midmark M11 - v1.0.5
Total Cycles: 2715

Name: RITTER2
Sterilizer ID: V2313228
08/01/2023 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 71.5 F 0.0
4:00 76.5 F 0.0
6:00 95.4 F 0.0
8:00 146.9 F 0.0
10:00 179.3 F 0.0
12:00 202.4 F 0.0
14:00 213.7 F 3.0
16:00 222.1 F 6.0
18:00 231.0 F 9.4
20:00 240.5 F 13.5
22:00 239.6 F 10.9
24:00 236.7 F 9.1
26:00 239.8 F 9.8
28:00 249.7 F 15.1
30:00 261.4 F 21.8
32:00 271.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.4 F 28.8
2:00 272.2 F 29.1
3:00 272.0 F 28.5
4:00 272.7 F 29.3

Min 271.6 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 196.0 F 0.0
10:00 202.9 F 0.0
15:00 197.0 F 0.0
20:00 194.7 F 0.0
25:00 193.7 F 0.0
30:00 197.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 32:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:10:43
CYCLE COMPLETE

08/01/2023 09:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____