
Midmark M11 - v1.0.0
Total Cycles: 4652

Name: RITTER 1
Sterilizer ID: V2178198
06/18/2024 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.7 F 0.0
2:00 129.7 F 1.5
4:00 147.4 F 4.6
6:00 171.1 F 8.5
8:00 203.7 F 13.2
10:00 217.4 F 5.8
12:00 230.5 F 10.4
14:00 230.7 F 7.2
16:00 242.4 F 12.1
18:00 239.2 F 10.0
20:00 252.0 F 16.3
22:00 263.1 F 22.7
24:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 28.8
2:00 271.4 F 28.3
3:00 272.6 F 28.9
4:00 271.6 F 28.2

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 217.1 F 0.0
10:00 205.3 F 0.0
15:00 206.2 F 0.0
20:00 236.8 F 0.0
25:00 238.7 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:33
CYCLE COMPLETE

06/18/2024 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____