

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5030

Name: RITTER 1  
Sterilizer ID: V2178198  
10/28/2024 02:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 83.8 F 0.0  
2:00 84.5 F 0.0  
4:00 110.2 F 0.2  
6:00 133.6 F 0.5  
8:00 163.0 F 1.5  
10:00 180.6 F 4.5  
12:00 203.1 F 7.4  
14:00 209.5 F 11.0  
16:00 222.6 F 5.5  
18:00 226.2 F 7.9  
20:00 236.3 F 12.1  
22:00 235.7 F 8.8  
24:00 242.7 F 11.3  
26:00 245.4 F 12.7  
28:00 256.6 F 18.6  
30:00 266.8 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.3 F 28.7  
2:00 272.2 F 28.4  
3:00 271.5 F 27.9  
4:00 271.6 F 28.3

Min 271.3 F 27.8  
Max 272.4 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.0 F 0.0  
5:00 195.6 F 0.0  
10:00 196.4 F 0.0  
15:00 194.8 F 0.0  
20:00 205.9 F 0.0  
25:00 225.9 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:42  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:10:21  
CYCLE COMPLETE

10/28/2024 03:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_