
Midmark M11 - v1.0.0
Total Cycles: 1384

Name: RITTER 1
Sterilizer ID: V2178198
11/18/2021 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.4 F 0.0
2:00 180.7 F 3.1
4:00 206.1 F 7.5
6:00 222.0 F 13.2
8:00 231.3 F 8.2
10:00 240.7 F 12.5
12:00 240.9 F 11.1
14:00 237.6 F 9.4
16:00 252.1 F 16.3
18:00 264.2 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 28.6
2:00 272.0 F 28.9
3:00 272.0 F 28.5
4:00 272.0 F 28.9

Min 271.2 F 28.2
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 228.5 F 0.0
10:00 214.5 F 0.0
15:00 215.6 F 0.0
20:00 206.1 F 0.0
25:00 206.3 F 0.0
Dry Cycle Aborted

FILLING: 1:54
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:10
DRYING: 28:48
TOTAL CYCLE: 00:57:46

11/18/2021 12:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____