

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 005091  
15:28 07/02/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:51  
PRESSURIZING 2:52  
STERILIZING 3:50  
136.4 C 222kPa 3:50  
136.4 C 225kPa 4:20  
136.3 C 224kPa 4:50  
136.5 C 221kPa 5:20  
136.4 C 225kPa 5:50  
136.2 C 221kPa 6:20  
136.4 C 219kPa 6:50  
136.4 C 225kPa 7:20  
136.4 C 225kPa 7:20  
VENTING 7:20  
AIR DRYING 8:17  
CYCLE COMPLETE 16:13

-----

Digital Signature #  
218ADD88DF92D6FB  
-----