
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002066
9:21 20/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:05
STERILIZING 3:07
136.0 C 222kPa 3:07
135.8 C 223kPa 3:37
136.3 C 226kPa 4:07
136.4 C 226kPa 4:37
136.3 C 224kPa 5:07
136.4 C 224kPa 5:37
136.4 C 227kPa 6:07
136.4 C 226kPa 6:37
136.4 C 226kPa 6:37
VENTING 6:37
AIR DRYING 7:27
CYCLE COMPLETE 11:50

Digital Signature #
134BBABD97CF56B4
