
Midmark M11 - v1.0.0
Total Cycles: 2357

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2022 08:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.6 F 0.0
2:00 143.4 F 1.4
4:00 144.5 F 4.5
6:00 173.3 F 8.8
8:00 197.8 F 10.7
10:00 213.7 F 6.1
12:00 227.8 F 10.7
14:00 229.6 F 7.1
16:00 239.7 F 12.1
18:00 237.8 F 9.9
20:00 250.3 F 16.2
22:00 261.5 F 22.6
24:00 272.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.7 F 29.9
2:00 272.3 F 29.3
3:00 271.8 F 29.2
4:00 272.7 F 29.9

Min 271.7 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 199.1 F 0.0
10:00 210.7 F 0.0
15:00 213.3 F 0.0
20:00 214.0 F 0.0
25:00 212.0 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:35
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:02:60
CYCLE COMPLETE

09/22/2022 09:31 AM
ACCEPTED: X
REJECTED:
Operator: U02