

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3318

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2024 04:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.6 F 0.0  
2:00 135.1 F 1.0  
4:00 157.8 F 3.9  
6:00 181.8 F 7.6  
8:00 201.7 F 11.8  
10:00 220.7 F 5.2  
12:00 228.9 F 9.4  
14:00 237.1 F 10.5  
16:00 240.7 F 11.4  
18:00 238.9 F 10.1  
20:00 252.0 F 16.4  
22:00 263.1 F 23.0  
24:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.7 F 29.2  
2:00 272.9 F 29.6  
3:00 271.5 F 29.1  
4:00 272.9 F 29.4

Min 271.5 F 28.9  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.1 F 0.0  
5:00 233.4 F 0.0  
10:00 212.4 F 0.0  
15:00 187.3 F 0.0  
20:00 184.6 F 0.0  
25:00 183.0 F 0.0  
30:00 181.3 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 24:54  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:03:43  
CYCLE COMPLETE

01/24/2024 05:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_