
Midmark M11 - v1.0.5
Total Cycles: 3721

Name: RITTER2
Sterilizer ID: V2313228
06/19/2024 05:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.3 F 0.0
2:00 116.0 F 1.1
4:00 136.0 F 3.9
6:00 169.4 F 7.5
8:00 188.7 F 11.8
10:00 213.0 F 5.2
12:00 223.3 F 9.9
14:00 230.7 F 8.6
16:00 238.7 F 11.8
18:00 238.0 F 10.1
20:00 250.8 F 16.4
22:00 262.1 F 23.0
24:00 272.7 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 271.6 F 29.4
2:00 271.8 F 29.4
3:00 272.2 F 29.6
4:00 272.5 F 29.8

Min 271.5 F 29.4
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 229.2 F 0.0
10:00 197.9 F 0.0
15:00 193.9 F 0.0
20:00 193.1 F 0.0
25:00 199.7 F 0.0
30:00 214.7 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 25:02
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:43
CYCLE COMPLETE

06/19/2024 06:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____