

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3444

Name: RITTER2  
Sterilizer ID: V2313228  
03/08/2024 10:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 70.5 F 0.0  
4:00 76.7 F 0.0  
6:00 102.4 F 0.3  
8:00 140.9 F 2.6  
10:00 164.8 F 5.4  
12:00 180.0 F 8.7  
14:00 199.1 F 12.5  
16:00 211.9 F 4.8  
18:00 223.4 F 8.5  
20:00 234.1 F 12.7  
22:00 234.8 F 9.3  
24:00 239.6 F 10.6  
26:00 245.4 F 13.1  
28:00 256.3 F 19.0  
30:00 266.5 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.3 F 29.2  
2:00 271.5 F 29.0  
3:00 272.0 F 29.1  
4:00 271.4 F 29.0

Min 271.2 F 28.6  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 207.7 F 0.0  
10:00 191.4 F 0.0  
15:00 187.7 F 0.0  
20:00 184.5 F 0.0  
25:00 181.2 F 0.0  
30:00 178.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:40  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:09:53  
CYCLE COMPLETE

03/08/2024 11:14 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_