
Midmark M11 - v1.0.5
Total Cycles: 3770

Name: RITTER2
Sterilizer ID: V2313228
07/11/2024 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.4 F 0.0
2:00 105.0 F 0.4
4:00 113.2 F 2.3
6:00 141.9 F 5.1
8:00 166.3 F 8.3
10:00 184.7 F 11.9
12:00 211.1 F 4.8
14:00 221.2 F 8.6
16:00 231.5 F 12.8
18:00 232.1 F 9.2
20:00 238.8 F 11.6
22:00 242.0 F 12.3
24:00 253.3 F 17.8
26:00 262.8 F 23.7
28:00 272.2 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.6
1:00 272.8 F 30.8
2:00 272.1 F 30.1
3:00 272.2 F 30.8
4:00 272.5 F 30.3

Min 271.0 F 29.5
Max 273.4 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 199.8 F 0.0
10:00 201.1 F 0.0
15:00 199.4 F 0.0
20:00 200.6 F 0.0
25:00 200.9 F 0.0
30:00 195.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 28:30
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:07:20
CYCLE COMPLETE

07/11/2024 02:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____