
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000502
7:58 22/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:44
STERILIZING 4:55
136.0 C 220kPa 4:55
136.1 C 222kPa 5:25
136.0 C 220kPa 5:55
136.2 C 222kPa 6:25
136.4 C 223kPa 6:55
136.2 C 221kPa 7:25
136.3 C 223kPa 7:55
136.1 C 223kPa 8:25
136.1 C 223kPa 8:25
VENTING 8:25
AIR DRYING 9:19
CYCLE COMPLETE 9:28

Digital Signature #
8EEC9EEA81F686E6
