
Midmark M11 - v1.0.5
Total Cycles: 4124

Name: RITTER2
Sterilizer ID: V2313228
11/15/2024 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 75.2 F 0.0
4:00 111.3 F 0.0
6:00 108.5 F 0.2
8:00 148.9 F 0.3
10:00 175.1 F 0.4
12:00 196.8 F 1.0
14:00 208.2 F 3.9
16:00 217.6 F 6.9
18:00 227.3 F 10.4
20:00 231.7 F 8.9
22:00 238.1 F 10.8
24:00 235.6 F 8.4
26:00 242.7 F 11.3
28:00 246.7 F 13.7
30:00 258.3 F 20.1
32:00 268.8 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 29.1
2:00 271.9 F 29.0
3:00 272.3 F 28.9
4:00 272.0 F 29.1

Min 271.6 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 221.6 F 0.0
10:00 220.6 F 0.0
15:00 208.0 F 0.0
20:00 213.7 F 0.0
25:00 214.8 F 0.0
30:00 212.6 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:12
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:11:36
CYCLE COMPLETE

11/15/2024 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____