
Midmark M11 - v1.0.5
Total Cycles: 1172

Name: RITTER2
Sterilizer ID: V2313228
06/08/2022 08:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.6 F 0.0
2:00 145.6 F 1.3
4:00 159.4 F 4.2
6:00 184.6 F 8.4
8:00 203.7 F 13.4
10:00 218.3 F 7.9
12:00 233.8 F 12.9
14:00 234.8 F 9.9
16:00 235.1 F 8.9
18:00 247.0 F 14.5
20:00 258.9 F 21.0
22:00 269.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.2 F 30.5
2:00 272.1 F 30.3
3:00 272.0 F 29.9
4:00 272.5 F 30.0

Min 270.7 F 29.2
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 221.7 F 0.0
10:00 225.5 F 0.0
15:00 228.9 F 0.0
20:00 224.8 F 0.0
25:00 229.7 F 0.0
30:00 234.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 22:55
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

06/08/2022 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____