
Midmark M11 - v1.0.5
Total Cycles: 1491

Name: RITTER2
Sterilizer ID: V2313228
09/14/2022 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.7 F 0.0
2:00 188.7 F 3.9
4:00 196.0 F 9.1
6:00 218.1 F 8.4
8:00 230.7 F 10.8
10:00 232.6 F 8.5
12:00 239.1 F 10.2
14:00 247.4 F 14.6
16:00 260.3 F 21.7
18:00 271.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.4 F 30.0
2:00 272.2 F 29.5
3:00 271.7 F 29.5
4:00 272.8 F 30.0

Min 271.4 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 230.1 F 0.0
10:00 228.7 F 0.0
15:00 223.1 F 0.0
20:00 220.9 F 0.0
25:00 220.5 F 0.0
30:00 218.2 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 18:28
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

09/14/2022 02:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____