
Midmark M11 - v1.0.5
Total Cycles: 1992

Name: RITTER2
Sterilizer ID: V2313228
02/01/2023 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 218.0 F 0.0
2:00 211.9 F 3.9
4:00 218.6 F 9.8
6:00 221.6 F 6.3
8:00 237.0 F 11.7
10:00 236.4 F 9.7
12:00 236.3 F 8.9
14:00 251.0 F 16.1
16:00 263.5 F 23.4
18:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.4 F 29.4
2:00 271.5 F 28.9
3:00 272.5 F 29.5
4:00 271.7 F 28.8

Min 271.4 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 217.6 F 0.0
10:00 232.0 F 0.0
15:00 233.6 F 0.0
20:00 227.0 F 0.0
25:00 221.4 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 18:01
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:14
CYCLE COMPLETE

02/01/2023 11:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____