
Midmark M11 - v1.0.5
Total Cycles: 3408

Name: RITTER2
Sterilizer ID: V2313228
02/23/2024 02:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.1 F 0.0
2:00 126.5 F 0.6
4:00 141.7 F 3.0
6:00 171.6 F 6.4
8:00 191.3 F 10.1
10:00 208.9 F 7.8
12:00 221.8 F 7.2
14:00 232.7 F 11.5
16:00 230.6 F 8.3
18:00 242.7 F 12.7
20:00 242.9 F 12.5
22:00 254.5 F 18.5
24:00 264.8 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.3 F 29.9
2:00 271.3 F 29.2
3:00 272.4 F 30.4
4:00 272.1 F 29.9

Min 271.0 F 29.1
Max 273.1 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 217.8 F 0.0
10:00 215.8 F 0.0
15:00 207.1 F 0.0
20:00 211.3 F 0.0
25:00 215.5 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 25:59
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:04:43
CYCLE COMPLETE

02/23/2024 03:05 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____