

-----  
Midmark M11 - v1.0.5  
Total Cycles: 293

Name: RITTER2  
Sterilizer ID: V2313228  
08/19/2021 01:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.4 F 0.0  
2:00 182.5 F 3.9  
4:00 209.0 F 9.9  
6:00 223.8 F 6.9  
8:00 237.9 F 12.7  
10:00 239.8 F 11.4  
12:00 240.6 F 11.1  
14:00 255.2 F 18.5  
16:00 267.7 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 271.9 F 28.9  
2:00 272.3 F 29.4  
3:00 271.7 F 29.3  
4:00 271.8 F 29.0  
  
Min 271.4 F 28.7  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
5:00 220.2 F 0.0  
10:00 214.5 F 0.0  
15:00 208.4 F 0.0  
20:00 216.9 F 0.0  
25:00 234.3 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 3:57  
HEATING: 17:11  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:58:13  
CYCLE COMPLETE

08/19/2021 02:46 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02