

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000438  
12:19 14/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:39  
PRESSURIZING 3:40  
STERILIZING 4:37  
136.5 C 220kPa 4:37  
136.1 C 221kPa 5:07  
136.4 C 223kPa 5:37  
136.3 C 223kPa 6:07  
136.1 C 220kPa 6:37  
136.3 C 222kPa 7:07  
136.3 C 221kPa 7:37  
136.3 C 223kPa 8:07  
136.3 C 223kPa 8:07  
VENTING 8:07  
AIR DRYING 9:02  
CYCLE COMPLETE 9:05

-----  
Digital Signature #  
8D1C419A917D5DC9  
-----