

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2727

Name: RITTER 1  
Sterilizer ID: V2178198  
01/03/2023 10:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.6 F 0.0  
2:00 80.8 F 0.0  
4:00 116.0 F 0.0  
6:00 125.8 F 1.4  
8:00 152.1 F 3.2  
10:00 176.2 F 6.2  
12:00 194.7 F 10.0  
14:00 206.5 F 9.7  
16:00 217.6 F 5.6  
18:00 228.4 F 9.4  
20:00 238.3 F 13.5  
22:00 237.7 F 10.1  
24:00 237.0 F 8.7  
26:00 247.2 F 13.8  
28:00 259.7 F 20.5  
30:00 270.3 F 27.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.8 F 28.8  
2:00 271.5 F 28.1  
3:00 272.9 F 28.8  
4:00 271.7 F 28.0

Min 271.5 F 28.0  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 200.7 F 0.0  
10:00 206.1 F 0.0  
15:00 207.4 F 0.0  
20:00 208.2 F 0.0  
25:00 207.2 F 0.0  
30:00 192.1 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:52  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:08:44  
CYCLE COMPLETE

01/03/2023 11:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_