
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001923
9:07 05/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:37
STERILIZING 4:44
136.5 C 220kPa 4:44
136.2 C 223kPa 5:14
136.4 C 223kPa 5:44
136.4 C 222kPa 6:14
136.4 C 224kPa 6:44
136.3 C 222kPa 7:14
136.4 C 223kPa 7:44
136.3 C 224kPa 8:14
136.3 C 224kPa 8:14
VENTING 8:14
AIR DRYING 9:09
CYCLE COMPLETE 9:10

Digital Signature #
3DF04E4FA8859437
