
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000421
10:42 27/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:34
STERILIZING 3:40
136.0 C 228kPa 3:40
136.3 C 230kPa 4:10
136.4 C 232kPa 4:40
136.4 C 232kPa 5:10
136.1 C 230kPa 5:40
136.4 C 233kPa 6:10
136.3 C 232kPa 6:40
136.1 C 230kPa 7:10
136.1 C 230kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 8:27

Digital Signature #
F6EFA097E335E2C4
