

-----  
Midmark M11 - v1.0.5  
Total Cycles: 724

Name: RITTER2  
Sterilizer ID: V2313228  
01/25/2022 09:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.9 F 0.0  
2:00 79.2 F 0.0  
4:00 105.4 F 0.0  
6:00 116.4 F 0.8  
8:00 151.6 F 3.7  
10:00 177.1 F 6.7  
12:00 196.6 F 9.5  
14:00 208.2 F 13.4  
16:00 218.2 F 6.5  
18:00 229.3 F 10.5  
20:00 232.6 F 8.5  
22:00 240.8 F 12.1  
24:00 239.6 F 10.7  
26:00 251.8 F 16.7  
28:00 262.3 F 22.9  
30:00 272.3 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.1 F 29.6  
2:00 272.3 F 29.3  
3:00 271.5 F 29.2  
4:00 272.6 F 29.7

Min 271.4 F 28.8  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.5 F 0.0  
5:00 199.9 F 0.0  
10:00 199.8 F 0.0  
15:00 192.7 F 0.0  
20:00 188.2 F 0.0  
25:00 186.5 F 0.0  
30:00 185.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 30:21  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:08:39  
CYCLE COMPLETE

01/25/2022 10:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_