
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001025
12:18 21/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:47
STERILIZING 3:52
136.4 C 222kPa 3:52
136.1 C 223kPa 4:22
136.1 C 220kPa 4:52
136.4 C 222kPa 5:22
136.4 C 225kPa 5:52
136.1 C 223kPa 6:22
136.4 C 225kPa 6:52
136.3 C 224kPa 7:22
136.3 C 224kPa 7:22
VENTING 7:22
AIR DRYING 8:13
CYCLE COMPLETE 8:33

Digital Signature #
21145D40382BC202
