
Midmark M11 - v1.0.0
Total Cycles: 3758

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2023 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.1 F 0.0
2:00 176.8 F 0.3
4:00 196.9 F 4.1
6:00 211.2 F 9.3
8:00 225.5 F 9.2
10:00 232.2 F 9.5
12:00 237.2 F 9.5
14:00 244.6 F 12.9
16:00 243.3 F 11.8
18:00 256.1 F 18.4
20:00 267.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.7
2:00 271.6 F 28.0
3:00 272.6 F 28.9
4:00 272.2 F 28.4

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.0 F 0.0
5:00 219.7 F 0.0
10:00 228.1 F 0.0
15:00 230.4 F 0.0
20:00 224.2 F 0.0
25:00 219.9 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 21:27
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

09/12/2023 12:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____