
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002014
14:47 11/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:03
STERILIZING 3:06
136.4 C 223kPa 3:06
136.1 C 222kPa 3:36
136.5 C 226kPa 4:06
136.3 C 224kPa 4:36
136.1 C 222kPa 5:06
136.4 C 226kPa 5:36
136.3 C 225kPa 6:06
136.4 C 221kPa 6:36
136.4 C 221kPa 6:36
VENTING 6:36
AIR DRYING 7:30
CYCLE COMPLETE 19:30

Digital Signature #
4B8D92CF86A6E3B0
