
Midmark M11 - v1.0.5
Total Cycles: 838

Name: RITTER2
Sterilizer ID: V2313228
02/24/2022 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.1 F 0.0
2:00 169.3 F 3.4
4:00 184.1 F 8.2
6:00 208.7 F 9.8
8:00 228.3 F 9.1
10:00 236.3 F 9.4
12:00 244.0 F 13.2
14:00 244.2 F 12.5
16:00 257.7 F 19.6
18:00 269.0 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.8
1:00 272.7 F 29.7
2:00 272.7 F 29.8
3:00 272.1 F 29.8
4:00 271.8 F 29.3

Min 271.0 F 28.6
Max 273.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 229.4 F 0.0
10:00 218.5 F 0.0
15:00 210.9 F 0.0
20:00 202.3 F 0.0
25:00 195.3 F 0.0
30:00 194.8 F 0.0
Drying Complete

FILLING: 2:15
HEATING: 18:51
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

02/24/2022 02:48 PM
ACCEPTED: X
REJECTED:
Operator: U02