
Midmark M11 - v1.0.0
Total Cycles: 814

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2021 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.0 F 0.0
2:00 187.1 F 3.0
4:00 186.4 F 6.9
6:00 217.0 F 12.0
8:00 222.1 F 6.8
10:00 232.3 F 12.2
12:00 238.2 F 9.5
14:00 237.3 F 9.0
16:00 248.4 F 14.7
18:00 261.3 F 21.6
20:00 272.0 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 273.2 F 29.1
2:00 271.3 F 28.3
3:00 273.0 F 29.2
4:00 272.1 F 28.6

Min 271.1 F 28.3
Max 273.3 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 222.8 F 0.0
10:00 207.1 F 0.0
15:00 204.5 F 0.0
20:00 206.0 F 0.0
25:00 218.1 F 0.0
30:00 232.0 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

04/27/2021 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____