
Midmark M11 - v1.0.5
Total Cycles: 715

Name: RITTER2
Sterilizer ID: V2313228
01/20/2022 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.2 F 0.0
2:00 196.4 F 4.5
4:00 210.2 F 10.4
6:00 223.0 F 6.5
8:00 238.1 F 12.8
10:00 239.5 F 11.3
12:00 240.2 F 10.7
14:00 254.4 F 17.9
16:00 266.3 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.2
2:00 271.9 F 29.3
3:00 272.2 F 29.1
4:00 271.8 F 29.2

Min 271.4 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 208.9 F 0.0
10:00 218.7 F 0.0
15:00 214.6 F 0.0
20:00 213.9 F 0.0
25:00 213.9 F 0.0
30:00 212.0 F 0.0
Drying Complete

FILLING: 2:29
HEATING: 17:33
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:10
CYCLE COMPLETE

01/20/2022 02:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____