
Midmark M11 - v1.0.0
Total Cycles: 3639

Name: RITTER 1
Sterilizer ID: V2178198
08/11/2023 10:51 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 150.7 F 1.5
4:00 167.6 F 5.5
6:00 194.3 F 10.7
8:00 214.2 F 4.9
10:00 232.3 F 11.5
12:00 235.2 F 9.1
14:00 237.1 F 8.8
16:00 249.5 F 14.9
18:00 251.0 F 15.7
20:00 265.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.9 F 28.0
2:00 272.4 F 28.4
3:00 272.9 F 28.8

Min 271.6 F 27.9
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 228.0 F 0.0
Dry Cycle Aborted

FILLING: 1:01
HEATING: 21:29
STERILIZING: 3:00
VENTING: 2:59
DRYING: 1:05
TOTAL CYCLE: 00:29:34

08/11/2023 11:20 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____