
Midmark M11 - v1.0.0
Total Cycles: 2529

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2022 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.6 F 0.0
2:00 167.0 F 4.0
4:00 184.2 F 8.2
6:00 207.5 F 13.2
8:00 221.8 F 7.1
10:00 234.9 F 12.2
12:00 235.5 F 9.1
14:00 238.7 F 9.5
16:00 247.6 F 13.8
18:00 259.8 F 20.7
20:00 271.1 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.4 F 28.8
2:00 271.7 F 28.5
3:00 272.5 F 28.7
4:00 271.5 F 28.3

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 233.1 F 0.0
10:00 232.0 F 0.0
15:00 227.5 F 0.0
20:00 222.4 F 0.0
25:00 221.0 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

11/08/2022 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____