
Midmark M11 - v1.0.5
Total Cycles: 721

Name: RITTER2
Sterilizer ID: V2313228
01/24/2022 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.6 F 0.0
2:00 138.7 F 1.9
4:00 157.0 F 5.2
6:00 182.5 F 9.5
8:00 206.2 F 6.5
10:00 219.2 F 8.1
12:00 232.9 F 13.3
14:00 235.9 F 10.5
16:00 235.3 F 8.7
18:00 249.0 F 15.4
20:00 260.6 F 22.0
22:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.7 F 29.9
2:00 271.8 F 29.1
3:00 272.1 F 29.9
4:00 272.6 F 29.7

Min 271.4 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 196.7 F 0.0
10:00 201.5 F 0.0
15:00 201.7 F 0.0
20:00 196.2 F 0.0
25:00 197.8 F 0.0
30:00 201.9 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

01/24/2022 11:16 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____