
Midmark M11 - v1.0.5
Total Cycles: 1528

Name: RITTER2
Sterilizer ID: V2313228
09/22/2022 03:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.9 F 0.0
2:00 188.0 F 3.8
4:00 197.7 F 8.9
6:00 216.1 F 8.5
8:00 229.2 F 9.8
10:00 233.8 F 8.8
12:00 242.9 F 12.3
14:00 245.1 F 13.4
16:00 258.0 F 20.4
18:00 269.5 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 29.8
2:00 271.6 F 29.0
3:00 272.0 F 29.8
4:00 272.4 F 29.7

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.8 F 0.0
5:00 223.6 F 0.0
10:00 239.4 F 0.0
15:00 241.2 F 0.0
20:00 228.2 F 0.0
25:00 223.0 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 19:00
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

09/22/2022 04:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____