

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1415

Name: RITTER2  
Sterilizer ID: V2313228  
08/25/2022 11:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.7 F 0.0  
2:00 197.5 F 3.0  
4:00 202.9 F 8.0  
6:00 224.7 F 12.9  
8:00 229.8 F 9.7  
10:00 234.6 F 8.9  
12:00 244.2 F 12.7  
14:00 244.8 F 13.0  
16:00 257.7 F 19.9  
18:00 269.1 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.6 F 29.5  
2:00 271.4 F 28.7  
3:00 272.4 F 29.7  
4:00 272.2 F 29.2  
  
Min 271.2 F 28.7  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
5:00 218.3 F 0.0  
10:00 237.6 F 0.0  
15:00 228.9 F 0.0  
20:00 223.0 F 0.0  
25:00 220.6 F 0.0  
30:00 218.8 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 19:06  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:06  
CYCLE COMPLETE

08/25/2022 12:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_