

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002612  
10:59 03/11/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 2:49  
STERILIZING 3:54  
136.4 C 223kPa 3:54  
136.2 C 221kPa 4:25  
136.3 C 222kPa 4:55  
136.2 C 222kPa 5:25  
136.0 C 221kPa 5:55  
136.4 C 224kPa 6:25  
136.2 C 223kPa 6:55  
136.2 C 222kPa 7:25  
136.2 C 222kPa 7:25  
VENTING 7:25  
AIR DRYING 8:18  
CYCLE COMPLETE 8:20

-----  
Digital Signature #  
CFB1C5B26441B91B  
-----