
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006152
13:12 17/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:25
STERILIZING 3:33
136.4 C 221kPa 3:33
136.4 C 222kPa 4:03
136.2 C 222kPa 4:33
136.2 C 223kPa 5:03
136.4 C 225kPa 5:33
136.4 C 224kPa 6:03
136.4 C 224kPa 6:33
136.2 C 222kPa 7:03
136.2 C 222kPa 7:03
VENTING 7:03
AIR DRYING 7:55
CYCLE COMPLETE 8:00

Digital Signature #
325D7CE5E20D639B
