

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001252  
7:07 24/02/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 2:59  
STERILIZING 4:04  
136.4 C 223kPa 4:04  
136.3 C 224kPa 4:34  
136.4 C 222kPa 5:04  
136.3 C 219kPa 5:34  
136.3 C 223kPa 6:04  
136.3 C 223kPa 6:34  
136.4 C 221kPa 7:04  
135.9 C 219kPa 7:34  
135.9 C 219kPa 7:34  
VENTING 7:34  
AIR DRYING 8:25  
CYCLE COMPLETE 10:50

-----  
Digital Signature #  
5ED57F800AA374E3  
-----