
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001272
9:35 27/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:45
STERILIZING 5:24
136.4 C 223kPa 5:24
136.1 C 222kPa 5:54
136.2 C 223kPa 6:24
136.1 C 222kPa 6:54
136.4 C 223kPa 7:24
136.4 C 224kPa 7:54
136.3 C 223kPa 8:24
136.4 C 224kPa 8:54
136.4 C 224kPa 8:54
VENTING 8:54
AIR DRYING 9:51
CASSETTE REMOVED
CYCLE COMPLETE 11:49

Digital Signature #
752A52E673D6F689
