

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000518  
15:57 21/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:30  
PRESSURIZING 3:34  
STERILIZING 4:44  
136.0 C 229kPa 4:44  
136.4 C 233kPa 5:14  
136.4 C 233kPa 5:44  
136.4 C 230kPa 6:14  
136.3 C 233kPa 6:44  
136.4 C 233kPa 7:14  
136.4 C 233kPa 7:44  
136.3 C 231kPa 8:14  
136.3 C 231kPa 8:14  
VENTING 8:14  
CYCLE ABORTED 8:43  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 8:43  
-----

Digital Signature #  
7C6889A0A9819FEC  
-----