

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002885  
9:34 21/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:50  
PRESSURIZING 2:56  
STERILIZING 4:01  
136.0 C 220kPa 4:01  
136.4 C 223kPa 4:31  
136.3 C 222kPa 5:01  
136.2 C 220kPa 5:31  
136.1 C 222kPa 6:01  
136.3 C 222kPa 6:31  
136.4 C 222kPa 7:01  
136.4 C 222kPa 7:31  
136.4 C 222kPa 7:31  
VENTING 7:32  
AIR DRYING 8:24  
CYCLE COMPLETE 8:25

-----  
Digital Signature #  
3ED8601E147DCDC0  
-----