

-----  
Midmark M11 - v1.0.0  
Total Cycles: 726

Name: RITTER 1  
Sterilizer ID: V2178198  
03/17/2021 01:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.1 F 0.0  
2:00 159.3 F 2.2  
4:00 167.9 F 6.4  
6:00 196.8 F 11.7  
8:00 223.9 F 6.7  
10:00 227.9 F 12.9  
12:00 239.2 F 10.0  
14:00 232.7 F 7.8  
16:00 250.0 F 15.0  
18:00 262.3 F 22.0  
20:00 271.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.6 F 29.4  
2:00 271.9 F 29.3  
3:00 271.4 F 29.0  
4:00 271.6 F 28.9

Min 270.9 F 28.7  
Max 273.1 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.2 F 0.0  
5:00 244.8 F 0.0  
10:00 224.2 F 0.0  
15:00 186.4 F 0.0  
20:00 183.1 F 0.0  
25:00 180.8 F 0.0  
30:00 222.4 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 20:49  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

03/17/2021 02:31 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01