

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002512  
14:10 11/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:55  
PRESSURIZING 2:57  
STERILIZING 4:07  
136.4 C 226kPa 4:07  
136.4 C 225kPa 4:37  
136.4 C 222kPa 5:07  
136.4 C 226kPa 5:37  
136.1 C 224kPa 6:07  
136.2 C 224kPa 6:37  
136.4 C 223kPa 7:07  
136.2 C 224kPa 7:37  
136.2 C 224kPa 7:37  
VENTING 7:37  
AIR DRYING 8:33  
CYCLE COMPLETE 8:41  
-----

Digital Signature #  
1DA0B8F22F0A150A  
-----