
Midmark M11 - v1.0.0
Total Cycles: 1367

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2021 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 176.4 F 2.1
4:00 195.0 F 6.1
6:00 210.5 F 11.0
8:00 221.2 F 5.2
10:00 235.6 F 10.6
12:00 230.7 F 7.2
14:00 244.3 F 13.0
16:00 241.4 F 11.2
18:00 254.1 F 17.6
20:00 265.4 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.7 F 29.2
2:00 271.7 F 28.4
3:00 272.7 F 29.3
4:00 272.6 F 28.9

Min 271.5 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.1 F 0.0
5:00 230.1 F 0.0
10:00 207.2 F 0.0
15:00 205.0 F 0.0
20:00 220.9 F 0.0
25:00 231.5 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:20
CYCLE COMPLETE

11/15/2021 10:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____