
Midmark M11 - v1.0.5
Total Cycles: 1718

Name: RITTER2
Sterilizer ID: V2313228
11/14/2022 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.5 F 0.0
2:00 187.0 F 2.0
4:00 200.1 F 7.0
6:00 223.7 F 12.6
8:00 229.3 F 8.8
10:00 236.4 F 10.3
12:00 243.5 F 13.0
14:00 243.7 F 12.3
16:00 256.8 F 19.3
18:00 268.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.4 F 29.3
2:00 271.5 F 28.9
3:00 272.5 F 29.3
4:00 271.5 F 28.7

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 217.9 F 0.0
10:00 233.3 F 0.0
15:00 229.9 F 0.0
20:00 228.3 F 0.0
25:00 227.4 F 0.0
30:00 225.6 F 0.0

Drying Complete

FILLING: 1:40
HEATING: 19:12
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

11/14/2022 03:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____