
Midmark M11 - v1.0.5
Total Cycles: 4281

Name: RITTER2
Sterilizer ID: V2313228
01/22/2025 04:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.8 F 0.0
2:00 142.7 F 1.6
4:00 155.2 F 4.7
6:00 182.6 F 8.8
8:00 200.4 F 13.1
10:00 216.5 F 6.1
12:00 230.6 F 10.7
14:00 232.8 F 7.6
16:00 243.1 F 12.8
18:00 242.0 F 11.4
20:00 254.1 F 17.8
22:00 265.1 F 24.4
24:00 270.7 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.0 F 29.7
2:00 272.0 F 29.8
3:00 271.9 F 29.7
4:00 271.9 F 29.6

Min 271.8 F 29.4
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 191.3 F 0.0
10:00 204.4 F 0.0
15:00 183.0 F 0.0
20:00 180.4 F 0.0
25:00 189.3 F 0.0
30:00 195.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:36
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:03:31
CYCLE COMPLETE

01/22/2025 05:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____