
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001232
10:15 29/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:49
STERILIZING 5:09
136.0 C 220kPa 5:09
136.4 C 224kPa 5:39
136.4 C 225kPa 6:09
136.0 C 222kPa 6:39
136.3 C 223kPa 7:09
136.1 C 220kPa 7:39
136.4 C 223kPa 8:09
136.4 C 222kPa 8:39
136.4 C 222kPa 8:39
VENTING 8:39
AIR DRYING 9:34
CYCLE COMPLETE 24:34

Digital Signature #
16F1A24C373A3C97
