
Midmark M11 - v1.0.5
Total Cycles: 164

Name: RITTER2
Sterilizer ID: V2313228
07/06/2021 06:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 74.5 F 0.0
4:00 76.8 F 0.1
6:00 93.8 F 0.5
8:00 117.1 F 3.9
10:00 147.3 F 7.5
12:00 176.4 F 11.2
14:00 199.4 F 4.8
16:00 209.7 F 7.9
18:00 222.0 F 12.0
20:00 230.0 F 8.6
22:00 238.2 F 11.8
24:00 243.0 F 12.5
26:00 254.2 F 18.5
28:00 264.3 F 24.8
30:00 270.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 30.3
1:00 272.6 F 30.5
2:00 272.6 F 30.4
3:00 272.4 F 30.5
4:00 272.1 F 30.3

Min 271.7 F 30.3
Max 272.8 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 229.7 F 0.0
10:00 226.3 F 0.0
15:00 226.3 F 0.0
20:00 225.8 F 0.0
25:00 225.2 F 0.0
Dry Cycle Aborted

FILLING: 0:48
HEATING: 30:32
STERILIZING: 4:00
VENTING: 3:07
DRYING: 29:45
TOTAL CYCLE: 01:08:13

07/06/2021 08:07 AM
ACCEPTED: X
REJECTED:
Operator: U01